



MAC™ Two-Lumen Central Venous Access for use with 7 - 8 Fr. Catheters MULTI-LUMEN/PSI KIT

SKU / Article #: AK-11142-D

Considerations

Not made with natural rubber latex.

Includes Lidocaine

±Precaution: Contains Phthalate: DEHP

Sales Quantity/Case: 10

MAC™ Two-Lumen Central Venous Access for use with 7 - 8 Fr. Catheters SKU / Article #: AK-11142-D


Quantity	Component Description
1	Two-Lumen Central Venous Access Device: 4" (10 cm) Radiopaque Polyurethane with 9 Fr. Distal Lumen, Integral Hemostasis Valve, Extension Line Clamps
1	Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
1	Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
1	Injection Needle: 25 Ga. x 1" (2.54 cm)
1	Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
1	Pressure Transduction Probe
1	Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
1	Injection Needle: 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
1	Syringe: 3 mL Luer-Lock
1	Tissue Dilator: 9 Fr. (3.0 mm) x 20.7 cm
1	5 mL 1% Lidocaine HCl Solution and Alcohol Prep
1	3 mL Applicator 2% CHG and 70% IPA ChloraPrep®1 One-Step Solution with Hi-Lite Orange™ Tint
1	Obturator: 8 Fr.

Representative photo(s): actual product specifications may vary from that shown in the photo

MAC Two-Lumen Central Venous Access for use with 7 - 8 Fr. Catheters SKU / Article #: AK-11142-D	
Quantity	Component Description
1	Cath-Gard® Catheter Contamination Shield: 80 cm with TwistLock™ Adapter
1	Dust Cap: Vented
2	Dust Cap: Non-Vented
1	SharpsAway® Disposal Cup
1	Drape: 24" x 36" (61 cm x 91 cm) with 4" (10 cm) fenestration
1	Drape: 24" x 36" (61 cm x 91 cm)
1	Scalpel: #11
1	Stopcock: 4-way High-Flow
2	Gauze Pad: 2" x 2" (5 cm x 5 cm)
5	Gauze Pad: 4" x 4" (10 cm x 10 cm)
1	Suture: 3-0 Braided Silk with Straight Needle
1	Tubing±: 20" (50 cm) Extension

¹ A registered trademark of CareFusion or one of its subsidiaries.

Documents



The products in this catalog may not be available in all countries. Please contact your local representative.

Indications, contraindications, warnings and instructions for use can be found in the Product Instructions for Use available under the Documents section on this page, where an IFU is available.