

Representative photo(s): actual product specifications may vary from that shown in the photo

Arrowgard Blue Plus® Three-Lumen CVC Kit

MULTI LUMEN CVC KIT: 3L 7 FR X 8 IN

SKU / Article #: ASK-45703-NOL

Highlights

Antimicrobial: Arrowgard Blue Plus*
Sharps Safety

Considerations

*Chlorhexidine
Not made with natural rubber latex.
‡MR Conditional

Sales Quantity/Case: 5

Arrowg+ard Blue Plus® Three-Lumen CVC Kit

SKU / Article #: ASK-45703-NOL






Quantity	Component Description
1	Arrowg+ard Blue Plus® ¹ Three-Lumen Catheter: 7 Fr. (2.5 mm OD) x 20 cm
1	Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
1	Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
1	Pressure Transduction Probe
1	Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
1	Injection Needle: 25 Ga. x 1" (2.54 cm) and 3 mL Luer-Lock Syringe
1	Injection Needle: 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
1	Tissue Dilator: 8.5 Fr. (2.8 mm) x 10.2 cm
1	3 mL Applicator 2% CHG and 70% IPA Chloraprep® ² One-Step Solution with Hi-Lite Orange™ Tint
1	SecondSite™ Adjustable Hub: Fastener
1	SecondSite™ Adjustable Hub: Catheter Clamp
1	SharpsAway® II Locking Disposal Cup
1	SharpsAway® Disposal Cup
1	Drape: 24" x 36" (61 cm x 91 cm) with 4" (10 cm) fenestration
1	Drape: 41" x 55" with 4" fenestration
1	Safety Scalpel: #11
1	Staple Anchoring Devices† ³
2	Gauze Pad: 2" x 2" (5 cm x 5 cm)
5	Gauze Pad: 4" x 4" (10 cm x 10 cm)
1	Suture: 3-0 Braided Silk with Straight Needle

¹ Licensed under US Patent No. 7,329,412.

² A registered trademark of CareFusion or one of its subsidiaries.

³ Licensed under US Patent No. RE37,814.

Documents



SCROLL TO VIEW MORE →

The products in this catalog may not be available in all countries. Please contact your local representative.

Indications, contraindications, warnings and instructions for use can be found in the Product Instructions for Use available under the Documents section on this page, where an IFU is available.