Teleflex®

ARROW®

Vascular Access Product Catalog



Representative photo(s): actual product specifications may vary from that shown in the photo

Pressure Injectable Arrowg+ard Blue Plus® Two-Lumen CVC

PI CVC KIT: 2-L 8 FR X 20 CM AGB

SKU / Article #: CDC-45802-XPN1A

Highlights

Antimicrobial: Arrowg+ard Blue Plus*
Sharps Safety

Considerations

*Chlorhexidine
Not made with natural rubber latex.
Includes Lidocaine
±Precaution: Contains Phthalate: DEHP

Sales Quantity/Case: 5

Pressure Injectable Arrowg+ard Blue Plus® Two-Lumen CVC
SKU / Article #: CDC-45802-XPN1A

Quantity	Component Description
1	Arrowg+ard Blue Plus®[1] Two-Lumen Catheter: 8 Fr. (2.8 mm OD) x 20 cm, Pressure Injectable
1	Spring-Wire Guide, Nitinol, Marked: .032" (0.81 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow® GlideWheel™ Wire Advancer
1	Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
1	Injection Needle: SafetyGlide™[2] 25 Ga. x 1" (2.54 cm)
1	Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
1	Pressure Transduction Probe
1	Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
1	Injection Needle: SafetyGlide™[2] 23 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
1	Syringe: 3 mL Luer-Lock
1	Tissue Dilator: 9 Fr. (3.0 mm) x 10.2 cm

Pressure Injectable Arrowg+ard Blue Plus Two-Lumen CVC
SKU / Article #: CDC-45802-XPN1A

Quantity	Component Description
2	10 mL Luer-Lock Pre-Filled Saline Syringe
1	5 mL 1% Lidocaine HCl Solution and Alcohol Prep
2	3 mL Applicator 2% CHG and 70% IPA ChloraPrep®[3] One-Step Solution with Hi-Lite Orange™ Tint
1	SecondSite™ Adjustable Hub: Fastener
1	SecondSite™ Adjustable Hub: Catheter Clamp
1	SharpsAway® II Locking Disposal Cup
1	SharpsAway® Disposal Cup
1	Maximal Barrier Drape™ with 4" fenestration
1	Towel
1	Needle Holder
1	Safety Scalpel: #11
1	Checklist/CLIP Sheet
1	Sterile Procedure Sign
1	Medication Label: 1% Lidocaine
2	Gauze Pad: 2" x 2" (5 cm x 5 cm)
5	Gauze Pad: 4" x 4" (10 cm x 10 cm)
1	Surgical Apparel: Impervious Gown
1	Dressing: I.V. Advanced 8.5 cm x 11.5 cm
1	Dressing: BIOPATCH®[4]
1	Transducer Cover: 14 cm x 147 cm
1	Surgical Apparel: Mask with Eye Shield
1	Surgical Apparel: Bouffant Cap
1	Suture: 3-0 Silk with Curved Needle
1	Tubing±: 20" (50 cm) Extension
1	HemoHopper® Fluid Receptacle

[1] Licensed under US Patent No. 7,329,412.
[2] A trademark of Becton, Dickinson and Company.
[3] A registered trademark of CareFusion or one of its subsidiaries.
[4] A registered trademark of Johnson & Johnson Corporation.

Documents


SafetyGlide™ Protected Needle

Bevel Up + Level Arm Up

Fig. 1

Fig. 2

Directions for use:

1. For SafetyGlide™ needle in accordance with ANSI/AAMI sterility standard.
2. Aseptically insert needle into vein with bevel up.
3. If appropriate for treatment, blood may be aspirated or administered now via the needle hub for transparent system or set.
4. In accordance with ANSI/AAMI standards, avoid engaging needle by uncapped hand by using the hands-free technique (i.e., discard the needle directly during the engaging process).
5. Secure needle and tubing using standard techniques.
6. Manipulate the safety mechanism into a non-activated position (without pushing activation arm) to ensure optimal, uninterrupted needle tip visibility and control.
7. Visually confirm that the device is now fully safe and activate the needle by a one-hand, 45° needle tip activation. Activation should be immediately clear, apparent through vibration.
8. Activation of the protective mechanism may cause a standard rebound of blood that may remain at the needle tip after injection.
9. After injection, discard the engaged syringe-safety unit in accordance with applicable regulations and institutional policy.

Warning: Needle contains residual medication. Handle with care to avoid injury.

Warning: Needle cannot be reused after the needle or all sheaths during use and disposal.

SafetyGlide™ Protected Needle

Bevel Up + Level Arm Up

Fig. 1

Fig. 2

Directions for use:

1. For SafetyGlide™ needle in accordance with ANSI/AAMI sterility standard.
2. Aseptically insert needle into vein with bevel up.
3. If appropriate for treatment, blood may be aspirated or administered now via the needle hub for transparent system or set.
4. In accordance with ANSI/AAMI standards, avoid engaging needle by uncapped hand by using the hands-free technique (i.e., discard the needle directly during the engaging process).
5. Secure needle and tubing using standard techniques.
6. Manipulate the safety mechanism into a non-activated position (without pushing activation arm) to ensure optimal, uninterrupted needle tip visibility and control.
7. Visually confirm that the device is now fully safe and activate the needle by a one-hand, 45° needle tip activation. Activation should be immediately clear, apparent through vibration.
8. Activation of the protective mechanism may cause a standard rebound of blood that may remain at the needle tip after injection.
9. After injection, discard the engaged syringe-safety unit in accordance with applicable regulations and institutional policy.

Warning: Needle contains residual medication. Handle with care to avoid injury.

Warning: Needle cannot be reused after the needle or all sheaths during use and disposal.




Vivacity™ 3.0 mL Prefilled Syringe with 0.5% Sodium Chloride Injection USP

Rx only

Sterile, Non-pyrogenic

Single Dose Prefilled Syringe

Directions for Use

Symbol	Title	Meaning	Standard	Reference
	Manufacturer	Indicates the medical device manufacturer.	ISO 15223-1:2016	5.1.1
	Use-by date	Indicates the date after which the medical device is not to be used.	ISO 15223-1:2016	5.1.4
	Batch code	Indicates the manufacturer's batch code so that the batch or lot can be identified.	ISO 15223-1:2016	5.1.5
REF	Catalogue number	Indicates the manufacturer's catalogue number so that the medical device can be identified.	ISO 15223-1:2016	5.1.6
	Sterilized using irradiation	Indicates a medical device that has been sterilized using irradiation.	ISO 15223-1:2016	5.2.4
	Do not resterilize	Indicates a medical device that is not to be resterilized.	ISO 15223-1:2016	5.2.7
	Do not reuse	Indicates a medical device that is intended for use only on a single procedure and then is to be discarded.	ISO 15223-1:2016	5.2.8
	Caution	Indicates the need for the user to consult the instructions for use for important cautionary information such as warnings and precautions that cannot, for various reasons, be presented on the medical device itself.	ISO 15223-1:2016	5.4.4
Rx only	Prescription device	Caution: Federal (USA) law restricts this device to sale by or on the order of a physician.	21 CFR 801.109	

Vivacity 3.0 mL Prefilled Syringe with 0.5% Sodium Chloride Injection USP

Rx only

Sterile, Non-pyrogenic

Single Dose Prefilled Syringe

Manufactured by:
Medefil, Inc.
1638 Gateway Drive
San Marcos, CA 92078 USA

OPEN HERE

Lidocaine for use in Vivacity™ 3.0 mL Apply alcohol wipe to rubber stopper. Hold stopper before removing cap. Twist/pull cap to open. Follow your facility's safety guidelines.

OPEN HERE

Lidocaine for use in Vivacity™ 3.0 mL Apply alcohol wipe to rubber stopper. Hold stopper before removing cap. Twist/pull cap to open. Follow your facility's safety guidelines.

OPEN HERE

Lidocaine for use in Vivacity™ 3.0 mL Apply alcohol wipe to rubber stopper. Hold stopper before removing cap. Twist/pull cap to open. Follow your facility's safety guidelines.

OPEN HERE

Lidocaine for use in Vivacity™ 3.0 mL Apply alcohol wipe to rubber stopper. Hold stopper before removing cap. Twist/pull cap to open. Follow your facility's safety guidelines.


ARROW

Tegaderm™ I.V. Advanced Securement Dressing

Rx only

Product Description:

Indications for Use:

Intended Purpose:

Contraindications:

Warnings:

Precautions:

Directions for Use:

Teleflex


ARROW

Pressure Injectable Arrowg+ard Blue Plus® Central Venous Catheter (CVC) Product

Rx only

Arrowg+ard Blue Plus® Antimicrobial Catheter Technology Information

Introduction

Rationale for Indwelling Catheters

Product Description

Indications for Use

Special Patient Populations

Pharmacokinetic Potential

Intended Use

Clinical Considerations

Clinical Evaluations:
Clinical Studies

Adverse Events

Indications for Use