

## Pressure Injectable Arrowg+ard Blue Advance® Three-Lumen PICC with 80 cm Hydrophilic-Coated Nitinol Guidewire

AGBA PICC SET: 3L 6FR X 55CM W/80CM SWG

SKU / Article #: PR-45563-HPHNM

### Highlights

Antimicrobial: Arrowg+ard Blue Advance\*

Antithrombogenic: Arrowg+ard Blue Advance\*

### Considerations

\*Chlorhexidine

Not made with natural rubber latex.

##MR Safe

**Sales Quantity/Case: 5**

*Representative photo(s): actual product specifications may vary from that shown in the photo*

# Pressure Injectable Arrowgard Blue Advance® Three-Lumen PICC with 80 cm Hydrophilic-Coated Nitinol Guidewire





SKU / Article #: PR-45563-HPHNM

Quantity	Component Description
1	Three Lumen TaperFree® Catheter with Arrowgard Blue Advance® Antimicrobial/Antithrombogenic Protection <sup>1</sup> : 6 Fr. (2.11 mm OD) x 55 cm, Pressure Injectable, Blue FlexTip®
1	GlideThru™ Peel-Away Sheath: 6 Fr. x 4" (10 cm) Radiopaque over 6 Fr. Dilator
1	Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other)
1	Guidewire, Marked: .018" (0.46 mm) dia. x 80 cm Hydrophilic-Coated Nitinol with Soft Tip Tungsten Coil (CE 0120)
1	Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
1	Syringe: 10 mL Luer-Lock
3	Dust Cap: Non-Vented
1	SecondSite™ Adjustable Hub: Fastener
1	SecondSite™ Adjustable Hub: Catheter Clamp
1	Catheter Trimmer
1	Safety Scalpel: #11
1	Patient ID Card
1	Chart Sticker
1	Patient Information Booklet
1	Checklist/CLIP Sheet
1	Paper Tape Measure
1	Dressing: STATLOCK® <sup>2</sup> PICC Plus Catheter Stabilization Device

<sup>1</sup> Licensed under US Patent No. 7,329,412.

<sup>2</sup> A registered trademark of C. R. Bard, Inc.

## Documents



SCROLL TO VIEW MORE →

The products in this catalog may not be available in all countries. Please contact your local representative.

Indications, contraindications, warnings and instructions for use can be found in the Product Instructions for Use available under the Documents section on this page, where an IFU is available.